



National Institute of Immunology
New Delhi-110067, INDIA

Notification for Inviting proprietary purchase

Dated: 23.07.2024

NII is planning the procurement of the following item through proprietary mode.

Name of the item: SW41Ti Rotor

Make:- Beckman Coulter

Model:-Optima XPN-100

Any other firm with item meeting the below mentioned specification may contact to equipment@nii.ac.in latest by 6-8-2024 or send the quotation in two bid system (technical and price separately)

Dated: 12th July 2024

Specification for SW41Ti rotor

for Ultra centrifuge (Optima XPN-100), Beckman Coulter

Swinging bucket, capacity 6 x 13.2 mL rotor,

Maximum speed: 41,000 rpm, 288,000 x g

Should be supplied with the following accessories:

- (1) 13.2mL, 14mm X 89mm, thin wall, Polypropylene (pack of 50)
- (2) 14mL, 14mm X 95mm, Ultra-Clear (pack of 50)
- (3) Hinge pin tool
- (4) Bucket holder rack
- (5) O-ring, Bucket
- (1) Over speed disk, 41,000 rpm
- (1) Spinkote lubricant
- (1) Vacuum grease